



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

CONCURRENCE 0-0-0-0000


ORIGINAL  
(Red)

Environmental Monitoring and  
Support Laboratory-Cincinnati

SUBJ: Analysis of Well Water Samples  
FROM: James Eichelberger and William L. Budde  
THRU: Dwight G. Ballinger  
Director, EMSL-Cincinnati  
TO: Orterio Villa  
Director,, Annapolis Field Office  
EPA, Region III

May 27, 1977

This will confirm my telephone report of May 20, 1977 concerning the well water samples taken near the Llangollen land fill. The samples were analyzed by the generalized GC/MS inert gas purge and trap procedure as described in chapter 5 of the draft GC/MS procedural manual. A commercial (Tekmar) device was used with a 25 ml sample volume. Detection limits were established for the compounds on your list in the range of 1-10  $\mu\text{g/l}$  using standards made up in this laboratory. No compounds were observed in any of the samples with this procedure.



RECEIVED

JUN - 7 1977

RECEIVED  
EPA, REGION III

AR000087